

Video Article

Measuring the Induced Membrane Voltage with Di-8-ANEPPS

Gorazd Pucihar, Tadej Kotnik, Damijan Miklavčič

Department of Biomedical Engineering, Faculty of Electrical Engineering, University of Ljubljana

Correspondence to: Gorazd Pucihar at gorazd.pucihar@fe.uni-lj.siURL: <http://www.jove.com/index/Details.stp?ID=1659>

DOI: 10.3791/1659

Citation: Pucihar G., Kotnik T., Miklavčič D. (2009). Measuring the Induced Membrane Voltage with Di-8-ANEPPS. JoVE. 33. <http://www.jove.com/index/Details.stp?ID=1659>, doi: 10.3791/1659

Abstract

Placement of a cell into an external electric field causes a local charge redistribution inside and outside of the cell in the vicinity of the cell membrane, resulting in a voltage across the membrane. This voltage, termed the induced membrane voltage (also induced transmembrane voltage, or induced transmembrane potential difference) and denoted by $\Delta\Phi$, exists only as long as the external field is present. If the resting voltage is present on the membrane, the induced voltage superimposes (adds) onto it. By using one of the potentiometric fluorescent dyes, such as di-8-ANEPPS, it is possible to observe the variations of $\Delta\Phi$ on the cell membrane and to measure its value noninvasively. di-8-ANEPPS becomes strongly fluorescent when bound to the lipid bilayer of the cell membrane, with the change of the fluorescence intensity proportional to the change of $\Delta\Phi$. This video shows the protocol for measuring $\Delta\Phi$ using di-8-ANEPPS and also demonstrates the influence of cell shape on the amplitude and spatial distribution of $\Delta\Phi$.

Protocol

Part I: Preliminary steps

1. In this experiment Chinese hamster ovary cell line (CHO-K1) is used. Cells are plated in Lab-Tek II chambers (2 wells, 4 cm² each) (Nalge Nunc, Germany) at $\sim 0.7 \times 10^5$ cells/ml in HAM-F12 culture medium supplemented with 8% fetal calf serum, 0.15 mg/ml L-glutamine, 16 mg/ml gentamicin (all from Sigma-Aldrich, Steinheim, Germany), and 200 units/ml Crystacillin (Pliva, Zagreb, Croatia), and incubated in 5% CO₂ at 37°C. Alternatively, cells can also be plated on #1 glass cover slips (0.13 to 0.16 mm thick), coated with a cellular adhesive such as polylysine.
2. Incubate the cells in their culture medium. Incubation lasting 2 to 4 hours yields cells that are still roughly spherical, but firmly attached to the surface with a small part of their membrane. Alternatively, after 16 to 20 hours of incubation, cells are fully attached to the surface and have more complex shapes, but most of them are still not dividing.
3. Prepare a 10 mM stock solution of di-8-ANEPPS (Invitrogen, Eugene, Oregon, USA) by adding 843 μ l of DMSO (Sigma-Aldrich, Steinheim, Germany) to 5 mg of the dye in the original Invitrogen vial. The stock solution can be stored in a refrigerator at 4°C for several months. Before starting the experiments, warm up the the solution until the crystals of DMSO dissolve.
4. Some cell lines may require the use of pluronic to ease the dye incorporation into the cell membrane. Pluronic can be purchased in 20% stock solution in DMSO (F-127, Invitrogen, Eugene, Oregon, USA), or a stock solution of the same concentration can be prepared by dissolving pluronic in DMSO. Stock solution of pluronic can be stored at room temperature.

Part II: Loading the cells with di-8-ANEPPS

1. Mix 3 μ l of 10 mM di-8-ANEPPS and 2.5 μ l of 20% pluronic in 1 ml of the Spinner modification (calcium-depleted version) of the Minimum Essential Medium SMEM (medium M8167 or M4767, Sigma-Aldrich, Steinheim, Germany) in a 1.5 ml Eppendorf tube. This yields a "loading solution" containing approximately 30 μ M di-8-ANEPPS and 0.05% of pluronic. For other cell types, their native culture media can be used instead of SMEM.
2. Replace the culture medium in the Lab-Tek chamber with the loading solution. Transfer the chamber to the refrigerator for 10 min at 4°C. At this temperature, internalization of the dye through the plasma membrane is largely inhibited.
3. After staining, gently wash the excess dye two to three times with pure SMEM.
4. Leave 1.5 ml of SMEM in the chamber.

Part III: Experiment and image acquisition

1. The cells are observed using a fluorescence microscope (in our case Zeiss AxioVert 200, Zeiss, Germany) equipped with an oil immersion objective (x63, NA 1.4), a monochromator (Polychrome IV, Visitron, Germany) and a cooled CCD camera (VisiCam 1280, Visitron, Germany). The images are acquired with MetaFluor 7.1.1 and processed in MetaMorph 7.1.1 (both Molecular Devices, Downingtown, PA, USA), but other similar acquisition software can also be used. Set the excitation wavelength in the acquisition software to 490 nm and choose a suitable emission filter for ANEPPS, for example, a band-pass filter centered at 605 nm (605/55m, dichroic 565 DCXR, Chroma, Rockingham, VT, USA). If a monochromator is not available, an excitation filter centered at 490 nm can be used instead.
2. Place the chamber with cells onto the microscope stage, position the electrodes at the bottom of the chamber and connect them to the pulse generator. The medium should cover the electrodes.
3. To maintain cell viability and reduce the heating, the pulse should be of sufficiently low amplitude and short duration. In this experiment a square pulse with 35 V amplitude and 50 ms duration (low enough to avoid electroporation) is generated using a DC voltage supply and a custom-made microprocessor-controlled switcher device. Alternatively, a commercial pulse generator, such as 33210A (Agilent, Santa Clara, CA, USA) connected to an amplifier or a commercial stimulator, such as S88 (Grass, West Warwick, RI, USA), can be used. The pulse is delivered to two parallel Pt/Ir wire electrodes with 0.8 mm diameter and 4 mm distance between them, creating an electric field of

approximately 88 V/cm between the electrodes and inducing $\Delta\Phi$. The exact field distribution between the electrodes used in this experiment is described elsewhere [1].

4. Find the cells of interest. Apply a single electric pulse or a sequence of pulses. For each pulse, acquire two fluorescence images: one immediately before the pulse (the control image), and one during the pulse (the pulse image). Due to the low response of di-8-ANEPPS, the changes in the fluorescence are difficult to discern by the naked eye and become apparent only after processing. The pulse delivery must be synchronized with the image acquisition, which is a feature of the acquisition software. To avoid photobleaching and possible heating, the illumination can be limited to the duration of the pulse.

Part IV: Image processing and analysis

1. Open the images in MetaMorph software. For each pulse, subtract the background in both the control and the pulse image.
2. Choose a cell and set the region of interest so that it corresponds to the membrane. Measure the fluorescence intensities along this region in the control and pulse image and transfer the values to a spreadsheet.
3. For each pulse, subtract the control data from the pulse data, and divide the result by control data to obtain the relative fluorescence changes. If a sequence of pulses is applied, the values of relative fluorescence changes determined for each pulse can be averaged to get a more reliable measurement.
4. Transform the relative fluorescence changes into $\Delta\Phi$ using a calibration curve. A rough estimation of this curve can be obtained from the literature, but for higher accuracy, it has to be measured for each particular setup. For the cells and the setup used in this experiment a 6% decrease in fluorescence corresponds to a 100 mV increase in $\Delta\Phi$ [2].
5. Finally, plot the voltage as a function of the relative arc length in a graphing software package such as Excel (Microsoft Corp., Redmond, WA, USA), Sigma Plot (Systat Software Inc., San Jose, CA, USA), or Origin (OriginLab Corp., Northampton, MA, USA). The curve can also be smoothed using a suitable filter, such as the moving average filter.

Discussion

Measurements of the induced membrane (transmembrane) voltage, $\Delta\Phi$, can be important in various experimental settings, such as studies of voltage-gated membrane channels, action potential propagation, cardiac cell stimulation, or cell membrane electroporation [3, 4, 5, 6, 7]. With simple cell shapes, $\Delta\Phi$ can be calculated analytically. For example, for a spherical cell, $\Delta\Phi$ is given by the Schwan's equation, which states that the voltage is proportional to the field strength and cell size and follows the cosine function along the membrane [8, 9]. For more complicated cell shapes, $\Delta\Phi$ can deviate considerably from the cosine and must be determined either numerically, using a computer [2, 10, 11], or experimentally, using a potentiometric dye [12, 13, 14, 15].

One of the potentiometric dyes widely used for this purpose is di-8-ANEPPS (di-8-butyl-amino-naphthyl-ethylene-pyridinium-propyl-sulfonate), a fast dye with excitation and emission spectra dependent on the membrane voltage, which allows noninvasive observations of the variations of the $\Delta\Phi$ on the cell membrane and to measure its value. In this video, we show an experimental approach for determination of $\Delta\Phi$ by using di-8-ANEPPS.

The dye was developed by Professor Leslie Loew and colleagues [13, 14] at the University of Connecticut and belongs to the class of fast-response dyes. di-8-ANEPPS is nonfluorescent in water and becomes strongly fluorescent when it incorporates into the lipid bilayer of the cell membrane. A change in $\Delta\Phi$ results in a change of the intramolecular charge distribution and corresponding changes in the spectral profile and intensity of the dye's fluorescence. The fluorescence intensity of di-8-ANEPPS varies proportionally to the change of $\Delta\Phi$; the response of the dye is linear for voltages ranging from -280 mV to +250 mV [4, 16]. Relatively small changes in fluorescence of the dye, uneven membrane staining, and dye internalization make di-8-ANEPPS less suitable for absolute measurements of membrane voltage, e.g. its resting component, although such efforts were also reported [17]. It is, however, suitable for measuring larger changes in membrane voltage, such as the onset of induced membrane voltage in nonexcitable cells exposed to external electric fields [12, 13], or action potentials in excitable cells [4, 5]. Although not applied here, di-8-ANEPPS also allows determination of $\Delta\Phi$ by ratiometric measurements of fluorescence excitation [18] or emission [19], which increases the sensitivity of the response and reduces the abovementioned effects. As di-8-ANEPPS stains the membrane, it can also be used plainly as a membrane marker [2].

One of the drawbacks of the dye is that it is prone to photobleaching, so that prolonged exposures to the strong light should be avoided. Calibration of the dye is performed with either (i) potassium ionophore valinomycin and a set of different potassium concentrations in external medium [2,18], or (ii) patch-clamp in voltage clamp mode [17].

Finally, with the measurements of $\Delta\Phi$ on spherical cells and cells of more complex shapes, the video demonstrates the influence of the cell shape on the amplitude and spatial distribution of $\Delta\Phi$. Thus for spherical cells $\Delta\Phi$ is close to a cosine, in agreement with Schwan's equation, while for more complicated cell shapes the spatial distribution of $\Delta\Phi$ is more intricate [20]...

Acknowledgements

This work was supported by the Slovenian Research Agency with project Z2-9229 and program P2-0249. This video represents the supplementary material for the "Electroporation-based Technologies and Treatments" scientific workshop and postgraduate course, organized biannually by the Faculty of Electrical Engineering at the University of Ljubljana, Slovenia.

References

1. Mazères, S., Šel, D., Golzio, M., Pucihar, G., Tamzali, Y., Miklavčič, D. & Teissie, J. Non invasive contact electrodes for in vivo localized cutaneous electropulsation and associated drug and nucleic acid delivery. *J. Control. Release* 134, 125-131 (2009).
2. Pucihar, G., Kotnik, T., Valič, B. & Miklavčič, D. Numerical determination of transmembrane voltage induced on irregularly shaped cells. *Annals Biomed. Eng.* 34, 642-652 (2006).
3. Huang, C.J., Harootunian, A., Maher, M.P., Quan, C., Raj, C.D., McCormack, K., Numann, R., Negulescu, P.A. & Gonzalez, J.E. Characterization of voltage-gated sodium-channel blockers by electrical stimulation and fluorescence detection of membrane potential. *Nat. Biotechnol.* 24, 439-446 (2006).
4. Bedlack, R.S., Wei, M., Fox, S.H., Gross, E. & Loew, L.M. Distinct electric potentials in soma and neurite membranes. *Neuron* 13, 1187-1193 (1994).

5. Cheng, D.K.L., Tung, L. & Sobie, E.A. Nonuniform responses of transmembrane potential during electric field stimulation of single cardiac cells. *Am. J. Physiol.* 277, H351-H362 (1999).
6. Teissie, J., Eynard, N., Gabriel, B. & Rols, M.P. Electroporation of cell membranes. *Adv. Drug. Del. Rev.* 35, 3-19 (1999).
7. Sersa G., Miklavčič D. (2008). Electrochemotherapy of tumours. *JoVE*. 22. <http://www.jove.com/index/details.stp?ID=1038>, doi:10.3791/1038
8. Kotnik, T., Bobanovič, F. & Miklavčič, D. Sensitivity of transmembrane voltage induced by applied electric fields: a theoretical analysis. *Bioelectrochem. Bioenerg.* 43, 285-291 (1997).
9. Schwan, H.P. Electrical properties of tissue and cell suspensions. *Adv. Biol. Med. Phys.* 5, 147-209 (1957).
10. Gowrishankar, T.R. & Weaver, J.C. An approach to electrical modeling of single and multiple cells. *Proc. Natl. Acad. Sci. USA* 100, 3203-3208 (2003).
11. Joshi, R.P., Hu, Q. & Schoenbach, K.H. Modeling studies of cell response to ultrashort, high-intensity electric fields - Implications for intracellular manipulation. *IEEE Trans. Plasma Sci.* 32, 1677-1686 (2004).
12. Gross, D., Loew, L.M. & Webb, W. Optical imaging of cell membrane potential changes induced by applied electric fields. *Biophys. J.* 50, 339-348 (1986).
13. Fluhler, E., Burnham, V.G. & Loew, L.M. Spectra, membrane binding, and potentiometric responses of new charge shift probes. *Biochemistry* 24, 5749-5755 (1985).
14. Loew, L.M. Voltage-sensitive dyes: measurement of membrane potentials induced by DC and AC electric fields. *Bioelectromagnetics Suppl.* 1, 179-189 (1992).
15. Hibino, M., Shigemori, M., Itoh, H., Nagayama, K. & Kinoshita, Jr.K. Membrane conductance of an electroporated cell analyzed by submicrosecond imaging of transmembrane potential. *Biophys. J.* 59, 209-220 (1991).
16. Lojewska, Z., Franks, D.L., Ehrenberg, B. & Loew, L.M. Analysis of the effect of medium and membrane conductance on the amplitude and kinetics of membrane potentials induced by externally applied electric fields. *Biophys. J.* 56, 121-128 (1989).
17. Zhang, J., Davidson, R.M., Wei, M.D. & Loew, L.M. Membrane electric properties by combined patch clamp and fluorescence ratio imaging in single neurons. *Biophys. J.* 74, 48-53 (1998).
18. Montana, V., Farkas, D.L. & Loew, L.M. Dual-wavelength ratiometric fluorescence measurements of membrane potential. *Biochemistry* 28, 4536-4539 (1989).
19. Knisley, S.B., Justice, R.K., Kong, W. & Johnson, P.L. Ratiometry of transmembrane voltage-sensitive fluorescent dye emission in hearts. *Am. J. Physiol. Heart Circ. Physiol.* 279, H1421-H1433 (2000).
20. Pucihar, G., Miklavčič, D. & Kotnik, T. A time-dependent numerical model of transmembrane voltage induction and electroporation of irregularly shaped cells. *IEEE T. Biomed. Eng.* 56, 1491-1501 (2009).